Certificate of Calibration

Electronic Laboratories, Inc.

75 Aileron Court, Suite 6 Westminster, MD 21157 Office (410) 871-9900

Customer:

GBW Associates, LLC

3178 Cardinal Dr.

Westminster, MD 21157

Certificate Number:

1102422

Date Cal:

10/24/22

Due Cal:

10/24/23

Statement of Calibration: Electronic Laboratories, Inc. certifies that the calibration of this item was performed in accordance with our registered ISO 9001 QMS using documented and approved methods, and standard(s) that is traceable to the National Institute of Standards and Technology. The calibration has been accomplished on the instrument listed below by comparison with standards maintained by Electronic Laboratories, Inc. The accuracy and stability of all standards maintained by Electronic Laboratories, Inc. is traceable to national standards maintained by the National Institute of Standards and Technology in Washington, D.C. and Boulder, Colorado. A complete record of all work performed is maintained by Electronic Laboratories, Inc. and is available for inspection upon request. Our effective ISO 9001 QMS incorporates principles of ISO 17025, ANSI/NCSL Z540 and MIL-STD-45662A specifications.

Manufacture:

Akron Brass

Model #

200 PSI

Nomenclature:

Pressure Gage

Serial Number:

904181

Procedure:

ELI-CSI-2001

Asset#

N/A

Tolerance:

Manufactures

Condition:

Received In Tolerance Returned In Tolerance

STANDARDS USED FOR CALIBRATION

STD#	I
CID#4000	Vaisala Hum
CID#2000	Flul
CID#4101	1

DESCRIPTION
Vaisala Humidity/Temperature Indicator
Fluke Pressure Module
Fluke Calibrator

04/23 06/23 06/23

CALIBRATED AT: 70 Deg °F / HUMIDITY AT: 49 %

Uncertainty factor does not exceed 25% of the tolerance of the characteristics being calibrated unless otherwise specified.

Cal'd - Certified by:	gover Russenburg	Title:	Calibration Engineer	Date:	10/24/22
	Josh Rubenking				

This Certificate is not a contract. We are not responsible for any damage resulting from the use of these standards by the customer. This certificate shall not be reproduced, except in full, without the written approval of ELI.